

1. Who is the incumbent provider of these products/services, if any?  
**No current contract, we buy through quoting process**
2. What does your agency currently pay for the products/services described in this bid?  
**That's confidential**
3. What is the County's current percent positive rate?  
**We consider this to be an open ended questions and we're not sure of its relevance to the RFP**
4. Will the County accept NON-CLIA waived Devices? CLIA-waived devices are typically used for facilities with active waived laboratory licenses through CLIA. Further, if an agency performs waived tests, CLIA requires that they follow the manufacturer's instructions for the waived tests, which includes performing quality control measures using positive and negative controls. Unless each County facility using these devices has a CLIA-waived license and intends to conduct quality control testing, CLIA waived products are not necessary.  
**CLIA-waived is preferred. We follow all manufacturer's instructions relative to the test cups we currently use and any presumptive positive screen are sent to Alere Labs as SAMSHA certified lab at discretion of the officer.**
5. Is the County flexible regarding what specimen validity (also called adulteration) measures are tested on the cup? For example, could the cup contain tests for creatinine, nitrites, pH, and specific gravity but not oxidants?  
**Yes**
6. Does the County require adulteration testing to be built into the cup, or would the County accept a separate strip for adulteration testing?  
**Build into cup**
7. If the County utilizes the confirmation option of this bid, will the County be sending in all presumptive positive tests or only select tests?  
**Discretion of county to send for confirmation**